

AB Tip İnceleme Sertifikası EU Type-Examination Certificate

Belge No / Certificate No : 124-20-01
Belgelendirme Tarihi - Bir Sonraki Belge Tarihi /
Certification Date / Certificate Validity Date : 15.01.2021-15.01.2026
Belge Geçerlilik Tarihi / Document Validity Period : 5 yıl / 5 years
Firma Unvanı ve Adresi /
Company Name and Address : INTEXTRED LTD.
4400 Pazardjik, Plovdivsko shose 107 str., Bulgaria

Ürün Adı /Modeller / Product Name / Models : 003
Direktifi / Directive : 2016/425 REGULATION
Modülü/Kategori / Module / Category : B MODÜLÜ/ KATEGORİ III
MODULE B / CATEGORY III
Test Rapor No/ları / Test Report No : M-2020-00772

Ürün Tipi / Product Type:
- EN 149:2001+ A1:2009 Solunumla ilgili koruyucu cihazlar - Parçacıklara karşı koruma amaçlı
filtreli yarım maskeler/ Respiratory protective devices - Filtering half masks to protect against
particles

Ürünün Malzeme Bilgisi / Product Material Information: 003 model ürünleri kumaş, kulak kayışı, burun
klipsi ve filtre katmanı kullanılarak imal edilmiştir./ 003 model products are manufactured using fabric, ear
loop, nose clip, filter layer.

Volkan AKIN
15.01.2021

Karar Verici / Approver

Okan AKEL
15.01.2021

Şirket Müdürü / General manager









ATTACHMENTS (124-20-01)

To certify the PPE product at Category III level, C2 or D module is accompanied by applying one of the conformity assessment methods along with the EU Type Examination (Module B).


Model : 003

PPE SPECIFICATION	PERFORMANCE LEVELS
Classification	FFP2
Reusable / Single Shift Use	NR

PPE produced as a single unit to fit an individual user, all the necessary instructions for manufacturing such PPE on the basis of the approved basic model:

MARKING	
MANUFACTURER: INTEXTRED LTD.	
PPE TYPE :	
- EN 149:2001+ A1:2009 Respiratory protective devices - Filtering half masks to protect against particles	
MODEL: 003	
PICTOGRAM AND PERFORMANCE LEVELS:	
EN 149:2001+ A1:2009 FFP2 NR	
 NB 2841	
 Year Month	 yyyy/mm
 -xx°C +yy°C	 < xx%
Or Condition of Storage	

MNA LABORATORIES SAN. TIC. LTD. ŞTİ declares that the above-mentioned product meets the requirements of the directive according to the EU Directive 2016/425, the safety of the product is covered by the conditions and use specified in this certificate and in the technical file.

ATTACHMENTS (124-20-01)**DOCUMENTS IN THE TECHNICAL FILE**

- Basic Health Safety Requirements
- Risk Assessment
- Test Reports
- Technical Report